

CERTIFICATE OF ANALYSIS
DEPLETED SERUM/PLASMA

Vitamin D Depleted Processed Serum

Code No.	SF504-7
Lot No.	N/A – Generic Example

RECEIVER INFORMATION

Storage Temperature	Store below -15°C.
Storage Notes	Avoid freeze thaw cycles.
Shipping Notes	Dry ice recommended.

PRODUCT INFORMATION

Source	Normal human serum from delipidated, defibrinated human plasma.
Presentation	Single homogenous batch, vitamin D2 and D3 depleted, filtered through 0.2µm membrane.

HEALTH AND SAFETY

Application	For Research and Manufacturing Only.
--------------------	--------------------------------------

Infectious Disease Tests

Starting material serum donor tested using FDA approved tests for:	Result
HIV 1 & 2 antibodies	Negative
Hepatitis B Surface Antigen	Negative
Hepatitis C virus antibodies	Negative

Precaution	No test can guarantee the absence of an infectious agent. Please handle as potentially hazardous.
-------------------	---

Material Safety	The product is supplied with safety data sheet HS0353. For further information and technical details, please request a Material Safety Data Sheet.
------------------------	--

ANALYSIS

ANALYTE	SPECIFICATION	RESULTS
25 OH Vitamin D3 and D2 level as determined by LC-MS	Report result.	<10 nmol/L.
Bioburden	<10 CFU/ml.	Pass.

QUALITY ASSURANCE

Manufactured under ISO 9001

Name:		Position:			
Signed:		Date:	Order #		

* Uncontrolled example document. Specifications are liable to change.